

## MANUAL PEDIATRIC BED

### TECHNICAL SPECIFICATION DATA SHEET



- Epoxy powder painted steel struction and corrosion resistant
- Manually performs the following positions:
  - Trendelenburg, Reverse Trendelenburg, Fowler, Leg elevation.
- Lateral handrails with adjustable fastening system, minimum of three positions or heights.
- Distance between rods no greater than 6 cm
- Telescopic lectern system
- Coupling point for lectern, Height from the floor to the top of the mattress minimum of 75 cm.
- Sleeping area usable dimensions 75cm width x 139 cm length  $\pm$ 15 cm
- Foam mattress with a cover of waterproof, anti-shell (anti-tear) material, tolerant to hospital detergents hospital-grade detergent tolerant, fire retardant and without aerators, not less than 10 cm thick.
- Mattress has handles on both sides for turning it over. Mattress dimension is compatible with the sleeping area
- Four (4) thermoplastic or conductive nylon wheels, with ball bearings of 7.5 cm in diameter with two brakes